#### INTOX EC/IR II

### Quality Assurance Worksheet

Instrument Serial Number

010517

Worksheet Start Date

12/29/2014

Location

Fairfax County Adult Detention Center

Address

10520 Judicial Dr., Fairfax, VA 22030

DFS Technician

Heather Stanton

License No.

18910

Laboratory

✓ On-Site

Site Specification: No detrimental environmental conditions exist.

V

Instrument Barometer (mm HG)

754

Reference Barometer (mm HG)

Reference Barometer(RB)Serial # 009113

**RB** Calibration Due

4/28/2015

Measurement Assurance Check

Standard (sea level)

**PA Target** 

minimum

maximum

Sample 1

0.300 Precision

-0.003

0.003

Sample 2 Sample 3

sample min sample max 0.000 0.000

Standard (sea level)

0

PA Target

minimum

maximum

Sample 1

Precision

-0.003

0.003

Sample 2

0

sample min sample max 0.000

0.000

Sample 3

Dry gas standard Lot No. (with tank no.)

0.100

AG419702-49

Replaced dry gas standard (+O-ring)

Installed at Location

T Removed to DFS-Central

Supplies

Mouthpieces

Certificates of Analysis

Operator Worksheet

Other:



Notes:



nstrument	<b>Serial</b>	Number	
nstrument	Senai	Number	

permande mother of children	er destablishment	APPROVISION A	conctin	acture in first
	1000			1/2
Charge of the Book of	1000		1112	12 3
EA SPECIAL PROPERTY IN	Director and the second		~~	2.0

Certification Date



Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

Date

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level) 0.250 Precision 0	PA Target	minimum 0.000 sample min 0.000	maximum 0.000 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level) 0.150 Precision 0	PA Target	minimum 0.000 sample min 0.000	maximum 0.000 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level) 0.080 Precision 0	PA Target	minimum -0.003 sample min 0.000	maximum 0.003 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level) 0.020 Precision 0	PA Target	minimum -0.003 sample min 0.000	maximum 0,003 sample max 0.000	Sample 1 Sample 2 Sample 3

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

the wotath

Issuing Analyst Date 12/30/14

)10





# COMMONWEALTH OF VIRGINIA DEPARTMENT OF FORENSIC SCIENCE

## CERTIFICATE OF BLOOD ALCOHOL ANALYSIS AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED		NAME OF COURT		:
INSTRUMENT, TEST,		DFS		
		1 1		
BREATH ANALYSIS		* .		
SAMPLE EXAMINED AND TEST CONDUCTED BY		AGENCY		
CHANHON UPARUED M		DFS Central	tah	
STANTON, HEATHER, M DFS LICENSE NUMBER	LICENSE EXPIRES	DATE TEST CONDUCTED	Lan	
18910 TEST EQUIPMENT NUMBER	10/01/2016	12/29/2014		
TEST EQUIPMENT NUMBER			公司 (12)	
010517	· .	经产品的	是自然的性性的	16年6年2月
RESULTS: TIME SAMPLE TAKEN	1:51 EST			
SAMPLE'S ALCOHOL CONTENT $\frac{0}{1}$	.00	G	RAMS PER 210 LITERS	OF BREATH
	,			and the second s
ATTEST:				•
I CERTIFY THAT THE ABOVE IS AN ACCURATE FEQUIPMENT AND IN ACCORDANCE WITH THE				
CONDUCTED IN ACCORDANCE WITH T	HE DEPARTMENT'S	S SPECIFICATIONS; TH	HAT PRIOR TO ADMINI	ISTRATION OF
THE TEST THE ACCUSED WAS ADVISED OF HIS F MENT USED TO PERFORM THE BREATH TEST, A				
THIS DAY OF	, 20	<del>•</del>		
BREATH TEST OPERATOR		1		
BREATH TEST OF EXATOR				
☐ I HAVE BECEIVED A CODY OF THIS	CEDITIEIO ATE OE	ANIAIVOIC		2.5
☐ I HAVE RECEIVED A COPY OF THIS	CERTIFICATE OF	VIVALI 219	SUBJECT'S SIGNATURE	
$\square$ subject refused to sign for CC	NDV OF CERTIFICA	TE OF ANAIVSIS		·
- Jobject Refosed to Ston For CC	OF CENTIFICA	TE OF ANALISIS	OPERATOR'S SIGNAT	URE



## IntoxNet MIS Report Report Generated 30 Dec 2014 at 08:19



Test Results

	Instrument Seri	al Number 010	517
Test # 000817 Subject Test			
Test Location 1 Department of	Test Location 2 Forens	sic Science	Test Location 3
Test Date 29 Dec 2014	Test Time	11:44	System Check Passed
	Remote/Local	Local	•
Operator's Last Name STANTON	Operator's First Name	HEATHER	Operator's Middle Initial M
Agency DFS Central Lab	•	License	Number 18910
Card Serial Number 118910	Effective Date	10/01/2014	Expiration Date 10/01/2016
Subject's Last Name INSTRUMENT		Subject's I	First Name TEST
Subject's Middle Initial	Subject's Date of Bir	-	Subject's Sex Male
Driver's License Number	•	Driv	ver's License Expiration 00/00/0000
Driver's License State		Court Nam	ne DFS
End Date 29 Dec 2014 End Time 11	:52 Result Tim	e 11:51 Re	suit Date 29 Dec 2014 Result 0.00
Data Type DIAG	Sample Value	Pass	Sample Time 11:45
Data Type BLK	Sample Value	0.000	Sample Time 11:46
Data Type CHK	Sample Value	0.098	Sample Time 11:46
Data Type BLK	Sample Value	0.000	Sample Time 11:48
Data Type SUBJ	Sample Value	0.000	Sample Time 11:49
Data Type BLK	Sample Value	0.000	Sample Time 11:50
Data Type SUBJ	Sample Value	0.000	Sample Time 11:51
Data Type BLK	Sample Value	0.000	Sample Time 11:52
Standard Type Dry Gas Std			Standard Value 0.099
Standard Lot Number AG419702-49		Sta	andard Expiration Date 07/16/2016
Tank Pressure 899		Barome	tric Pressure 754 mmHg
Blow Sample Number 1 Blow Dui	ration 3.28 sec	Blow Volume	1815 cc End-of-Blow Time 11:49
Blow Sample Number 2 Blow Dur	ration 3.36 sec	Blow Volume	1750 cc End-of-Blow Time 11:51
Tomper Evident Stemp #5022	702		Tort Status Code 0

Test Status Success

Tamper Evident Stamp 55933703

Test Status Code 0